



# PINNACLE — ANALYTICS —

## Potency Results

Sample Name: *GMO Live Resin*

Client: Yam Ops LLC

Client Batch ID:

Pinnacle-Analytics.com

613 Market St

Medford OR 97504

P:(541)300-8217

**Sample ID:** rC-HS-633-G316

**Matrix:** Concentrate

**Prep Analyst:** Megan A.

**Analysis Method:** 0711014 H4 9-12-2025 #1.lcm

**Sampling Method:** N/A

**Reference Method:** JCB 2009: HPLC/DAD

**Analysis Batch:** 2-16-2026 H4 101, 604, 606, 661 Solids

Date Sampled: 2/16/2026

Date Reported: 2/16/2026

Client License: N/A


For R&D Purposes Only

<b>Total THC</b> (THCA*0.877+d9-THC)	66.2%
<b>Total CBD</b> (CBDA*0.877+CBD)	0.26%
<b>Moisture Content</b>	N/A

Cannabinoid	% Weight	mg/g
CBDV	<LOQ	<LOQ
CBDA	<LOQ	<LOQ
CBGA	0.905	9.05
CBG	<LOQ	<LOQ
CBD	0.26	2.6
THCV	<LOQ	<LOQ
CBN	<LOQ	<LOQ
d9-THC	0.271	2.71
d8-THC	<LOQ	<LOQ
CBC	<LOQ	<LOQ
THCA	75.2	752.0
<b>Total Cannabinoids</b>	<b>76.64</b>	<b>766.0</b>

\*ORELAP Accredited Analyte

Limit Of Quantitation: 0.14%, analyte not measured



These test results may not be altered or reproduced except in full without the permission of Pinnacle Analytics. These results were generated following the Oregon Administrative Rules and in accordance with the NELAP Institute under ORELAP License #4152  
Report generated by Flex\_Potency\_Rev1\_2-9-2025

Kris Ford, PhD  
Lab Director



# PINNACLE ANALYTICS

## Quality Control Results

**Analyst:** Megan A.

**Analysis Batch:** 2-16-2026 H4 101, 604, 606, 661 Solids

Pinnacle-Analytics.com  
613 Market St  
Medford OR 97504  
P:(541)300-8217

	Duplicate RPD		LCS % Recovery		Method Blank	
	HS-0-G301-b	Limit	C-SL-021626	Limits	C-SB-021626	Limit
<b>CBDA</b>	<LOQ%	30%	98.3%	90-110%	<LOQ/2	LOQ/2
<b>CBD</b>	5.09%	30%	104.0%	90-110%	<LOQ/2	LOQ/2
<b>d9-THC</b>	2.03%	10%	97.7%	90-110%	<LOQ/2	LOQ/2
<b>d8-THC</b>	<LOQ%	30%	99.8%	90-110%	<LOQ/2	LOQ/2
<b>THCA</b>	<LOQ%	30%	97.8%	90-110%	<LOQ/2	LOQ/2

RPD: Relative Percent Difference between unknown sample and its duplicate

LCS: Laboratory Control Sample with known concentration

Case Comments: There were no divergences from ordinary Quality Control procedures or SOPs.



These test results may not be altered or reproduced except in full without the permission of Pinnacle Analytics. These results were generated following the Oregon Administrative Rules and in accordance with the NELAP Institute under ORELAP License #4152  
Report generated by Flex\_Potency\_Rev1\_2-9-2025

Kris Ford, PhD  
Lab Director